


SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

<p>Pennsylvania Standards Laboratory 2221 Forster Street, Room G-44A Harrisburg, PA 17125-0001 Toni Nagy Phone: 717-346-8115 Fax: 717-346-3820 E-mail: tonagy@pa.gov</p>	<p>Fields of Calibration Dimensional Time and Frequency Mechanical</p> <p>This laboratory is compliant to ANSI/NCSLI Z540-1-1994; Part 1. (NVLAP Code: 20/A01)</p>
---	---

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
DIMENSIONAL			
SURVEYING RODS and TAPES (20/D13)			
Surveying Tapes	0 ft to 16 ft	0.0079 in	Bench Method
	0 ft to 30 ft	0.015 in	
	0 ft to 45 ft	0.022 in	
	0 ft to 60 ft	0.029 in	
	0 ft to 75 ft	0.036 in	
	0 ft to 84 ft	0.043 in	
	0 ft to 90 ft	0.043 in	
	0 ft to 100 ft	0.051 in	
	0 ft to 105 ft	0.051 in	
	0 ft to 120 ft	0.058 in	
	0 ft to 135 ft	0.065 in	
	0 ft to 150 ft	0.072 in	
	0 ft to 165 ft	0.079 in	
	0 ft to 180 ft	0.087 in	
	0 ft to 184 ft	0.094 in	
	0 ft to 195 ft	0.094 in	
0 ft to 200 ft	0.10 in		
TIME and FREQUENCY			
STOPWATCHES and TIMERS (20/F05)			
Stopwatches	0 h to 24 h	0.11 s	

2023-02-17 through 2024-03-31
Effective dates




For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
MECHANICAL			
FORCE (20/M06)			
Force Gauge	30 lbf	0.063 lbf	
	25 lbf	0.063 lbf	
	20 lbf	0.068 lbf	
	15 lbf	0.075 lbf	
	10 lbf	0.075 lbf	
	0 lbf to 5 lbf	0.061 lbf	
MASS DETERMINATION (20/M08)			
Metric	30 kg	23 mg	Echelon II
	25 kg	23 mg	
	20 kg	9.1 mg	
	10 kg	3.0 mg	
	5 kg	1.7 mg	
	3 kg	1.4 mg	
	2 kg	1.2 mg	
	1 kg	0.28 mg	
	500 g	0.15 mg	
	300 g	0.10 mg	
	200 g	86 µg	
	100 g	40 µg	
	50 g	25 µg	
	30 g	25 µg	
	20 g	20 µg	
	10 g	11 µg	
	5 g	10 µg	
	3 g	6.1 µg	
	2 g	5.9 µg	
	1 g	4.6 µg	
	500 mg	4.7 µg	
	300 mg	3.2 µg	
	200 mg	2.5 µg	
100 mg	2.5 µg		
50 mg	1.8 µg		
30 mg	1.6 µg		
20 mg	1.3 µg		


2023-02-17 through 2024-03-31
Effective dates


For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
Avoirdupois	10 mg	1.2 µg	Echelon II
	5 mg	1.2 µg	
	3 mg	1.2 µg	
	2 mg	3.9 µg	
	1 mg	3.4 µg	
	1000 lb	0.90 g	
	500 lb	0.68 g	
	200 lb	0.53 g	
	100 lb	75 mg	
	50 lb	21 mg	
	30 lb	18 mg	
	25 lb	15 mg	
	20 lb	12 mg	
	10 lb	3.9 mg	
	5 lb	1.8 mg	
	3 lb	1.7 mg	
	2 lb	0.37 mg	
	1 lb	0.17 mg	
	0.5 lb	100 µg	
	0.3 lb	81 µg	
	0.2 lb	44 µg	
	0.1 lb	28 µg	
	0.05 lb	25 µg	
	0.03 lb	16 µg	
	0.02 lb	13 µg	
	0.01 lb	4.9 µg	
	0.005 lb	4.1 µg	
	0.003 lb	2.3 µg	
	0.002 lb	2.1 µg	
	0.001 lb	3.3 µg	
	8 oz	100 µg	
	4 oz	130 µg	
2 oz	70 µg		
1 oz	35 µg		

2023-02-17 through 2024-03-31
Effective dates


For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
Metric	1/2 oz	36 µg	Echelon III
	1/4 oz	22 µg	
	1/8 oz	21 µg	
	1/16 oz	9.3 µg	
	1/32 oz	7.8 µg	
	2500 kg	44 g	
	1000 kg	17 g	
	750 kg	15 g	
	500 kg	6.1 g	
	250 kg	3.1 g	
	200 kg	2.5 g	
	100 kg	1.4 g	
	50 kg	600 mg	
	30 kg	360 mg	
	25 kg	310 mg	
	20 kg	250 mg	
	10 kg	140 mg	
	5 kg	31 mg	
	3 kg	19 mg	
	2 kg	13 mg	
	1 kg	7.8 mg	
	500 g	6.2 mg	
	300 g	2.4 mg	
	200 g	1.8 mg	
	100 g	1.2 mg	
	50 g	0.69 mg	
	30 g	0.50 mg	
	20 g	0.38 mg	
	10 g	0.27 mg	
	5 g	0.20 mg	
	3 g	0.16 mg	
	2 g	0.14 mg	
	1 g	0.064 mg	
500 mg	50 µg		
300 mg	39 µg		
200 mg	51 µg		



2023-02-17 through 2024-03-31

Effective dates

For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
Avoirdupois	100 mg	27 µg	Echelon III
	50 mg	24 µg	
	30 mg	21 µg	
	20 mg	18 µg	
	10 mg	16 µg	
	5 mg	14 µg	
	3 mg	13 µg	
	2 mg	12 µg	
	1 mg	12 µg	
	5000 lb	43 g	
	3000 lb	27 g	
	2500 lb	25 g	
	2000 lb	16 g	
	1000 lb	5.5 g	
	500 lb	2.9 g	
	250 lb	1.5 g	
	200 lb	1.3 g	
	100 lb	0.54 g	
	50 lb	0.28 g	
	30 lb	0.18 g	
	25 lb	0.15 g	
	20 lb	130 mg	
	10 lb	28 mg	
	5 lb	15 mg	
	4 lb	13 mg	
	3 lb	10 mg	
	2 lb	7.4 mg	
	1 lb	2.7 mg	
	0.5 lb	1.9 mg	
	0.3 lb	1.3 mg	
	0.2 lb	1.0 mg	
	0.1 lb	0.67 mg	
	0.05 lb	0.43 mg	
0.03 lb	0.31 mg		
0.02 lb	0.27 mg		
0.01 lb	0.18 mg		




2023-02-17 through 2024-03-31
Effective dates

For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
Troy Ounce	0.005 lb	0.15 mg	
	0.003 lb	0.14 mg	
	0.002 lb	0.068 mg	
	0.001 lb	0.055 mg	
	12 oz	2.4 mg	
	8 oz	1.9 mg	
	6 oz	1.6 mg	
	4 oz	1.2 mg	
	2 oz	0.73 mg	
	1 oz	0.49 mg	
	1/2 oz	0.32 mg	
	1/4 oz	0.23 mg	
	1/8 oz	0.17 mg	
	1/16 oz	0.14 mg	
	1/32 oz	0.075 mg	
	0.5 oz	0.32 mg	
	0.3 oz	0.25 mg	
	0.2 oz	0.23 mg	
	0.1 oz	0.19 mg	
	0.05 oz	0.16 mg	
	0.02 oz	0.14 mg	
	0.01 oz	0.14 mg	
	200 oz t	73 mg	
	100 oz t	37 mg	
	50 oz t	20 mg	
	20 oz t	9.7 mg	
	10 oz t	9.7 mg	
	5 oz t	6.2 mg	
	2 oz t	1.4 mg	
	1 oz t	0.74 mg	
	0.5 oz t	0.38 mg	
	0.25 oz t	0.23 mg	
0.2 oz t	0.22 mg		
0.1 oz t	0.19 mg		

2023-02-17 through 2024-03-31
Effective dates


For the National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES

NVLAP LAB CODE 200869-0

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
Grain	0.05 oz t	0.16 mg	
	0.025 oz t	0.15 mg	
	0.02 oz t	0.14 mg	
	0.01 oz t	0.13 mg	
	0.005 oz t	0.13 mg	
	1000 gr	1.6 mg	
	500 gr	0.79 mg	
	200 gr	0.34 mg	
	100 gr	0.24 mg	
	50 gr	0.20 mg	
	20 gr	0.16 mg	
	10 gr	100 µg	
	5 gr	77 µg	
	2 gr	59 µg	
	1 gr	52 µg	
Wheel Load Weighers	30 001 lb to 40 000 lb	80 lb	Echelon III Calibrated in Pairs
	20 001 lb to 30 000 lb	80 lb	
	10 001 lb to 20 000 lb	61 lb	
	0 lb to 10 000 lb	50 lb	
	15 001 lb to 20 000 lb	48 lb	Calibrated Singly
	10 001 lb to 15 000 lb	45 lb	
	5001 lb to 10 000 lb	38 lb	
	0 lb to 5000 lb	33 lb	
Weight Carts	6000 lb	140 g	Echelon III
	5500 lb	130 g	
	5000 lb	130 g	
	4500 lb	120 g	
	4000 lb	110 g	
	3500 lb	110 g	

2023-02-17 through 2024-03-31

Effective dates




For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks	
	3000 lb	100 g		
VOLUME and DENSITY (20/M12)				
Volume	1500 gal	45 in ³	Transfer Method	
	1000 gal	31 in ³		
	500 gal	16 in ³		
	200 gal	6.2 in ³		
	100 gal	3.1 in ³		
	60 gal	1.8 in ³		
	50 gal	1.5 in ³		
	25 gal	0.92 in ³		
	10 gal	0.41 in ³		
	5 gal	0.35 in ³		4 in neck
	5 gal	0.20 in ³		3 in neck
	5000 L	0.65 L		
	3000 L	0.39 L		
	2000 L	0.27 L		
	1000 L	0.13 L		
	500 L	62 mL		
	250 L	30 mL		
	200 L	30 mL		
	120 L	20 mL		
	100 L	16 mL		
	60 L	11 mL		
	50 L	9.4 mL		
	40 L	8.1 mL		
20 L	5.9 mL			
Large Volume	120 gal	1.8 in ³	Gravimetric Method Ball Valve Butterfly Valve	
	100 gal	1.7 in ³		
	100 gal	1.7 in ³		
	75 gal	1.6 in ³		
	60 gal	1.1 in ³		
	50 gal	1.0 in ³		Ball Valve
	50 gal	0.94 in ³		Butterfly Valve
25 gal	0.53 in ³	Ball Valve		


2023-02-17 through 2024-03-31
Effective dates


For the National Voluntary Laboratory Accreditation Program

CALIBRATION AND MEASUREMENT CAPABILITIES (CMC) ^{Notes 1,2}

Measured Parameter or Device Calibrated	Range	Expanded Uncertainty ^{Note 3}	Remarks
	25 gal	0.42 in ³	Butterfly Valve
	20 gal	0.41 in ³	
	15 gal	0.41 in ³	
	500 L	37 mL	
	250 L	18 mL	
	200 L	16 mL	
	100 L	8.7 mL	
	60 L	6.7 mL	
	50 L	6.7 mL	
	40 L	5.5 mL	
Bottom Drain Prover	10 gal	0.34 in ³	
	5 gal	0.23 in ³	
Slicker Standard	10 gal	0.16 in ³	
	5 gal	0.11 in ³	
	1 gal	0.028 in ³	
Test Measure	10 gal	0.31 in ³	
	5 gal	0.15 in ³	
	5 gal (Imperial)	0.18 in ³	
Slicker Standard	20 L	1.8 mL	
	10 L	0.95 mL	
	5 L	0.50 mL	
Test Measure	20 L	2.6 mL	
END			

2023-02-17 through 2024-03-31
Effective dates


For the National Voluntary Laboratory Accreditation Program

Notes

Note 1: A Calibration and Measurement Capability (CMC) is a description of the best result of a calibration or measurement (result with the smallest uncertainty of measurement) that is available to the laboratory's customers under normal conditions, when performing more or less routine calibrations of nearly ideal measurement standards or instruments. The CMC is described in the laboratory's scope of accreditation by: the measurement parameter/device being calibrated, the measurement range, the uncertainty associated with that range (see note 3), and remarks on additional parameters, if applicable.

Note 2: Calibration and Measurement Capabilities are traceable to the national measurement standards of the U.S. or to the national measurement standards of other countries and are thus traceable to the internationally accepted representation of the appropriate SI (Système International) unit.

Note 3: The uncertainty associated with a measurement in a CMC is an expanded uncertainty with a level of confidence of approximately 95 %, typically using a coverage factor of $k = 2$. However, laboratories may report a coverage factor different than $k = 2$ to achieve the 95 % level of confidence. Units for the measurand and its uncertainty are to match. Exceptions to this occur when marketplace practice employs mixed units, such as when the artifact to be measured is labeled in non-SI units and the uncertainty is given in SI units (Example: 5 lb weight with uncertainty given in mg).

Note 3a: The uncertainty of a specific calibration by the laboratory may be greater than the uncertainty in the CMC due to the condition and behavior of the customer's device and specific circumstances of the calibration. The uncertainties quoted do not include possible effects on the calibrated device of transportation, long term stability, or intended use.

Note 3b: As the CMC represents the best measurement results achievable under normal conditions, the accredited calibration laboratory shall not report smaller uncertainty of measurement than that given in a CMC for calibrations or measurements covered by that CMC.

Note 3c: As described in Note 1, CMCs cover calibrations and measurements that are available to the laboratory's customers under *normal conditions*. However, the laboratory may have the capability to offer special tests, employing special conditions, which yield calibration or measurement results with lower uncertainties. Such special tests are not covered by the CMCs and are outside the laboratory's scope of accreditation. In this case, NVLAP requirements for the labeling, on calibration reports, of results outside the laboratory's scope of accreditation apply. These requirements are set out in Annex A.5 of NIST Handbook 150, Procedures and General Requirements.

Note 4: Uncertainties associated with field service calibration may be greater as they incorporate on-site environmental contributions, transportation effects, or other factors that affect the measurements. (This note applies only if marked in the body of the scope.)

Note 5: Values listed with percent (%) are percent of reading or generated value unless otherwise noted.

Note 6: NVLAP accreditation is the formal recognition of specific calibration capabilities. Neither NVLAP nor NIST guarantee the accuracy of individual calibrations made by accredited laboratories.

2023-02-17 through 2024-03-31

Effective dates



For the National Voluntary Laboratory Accreditation Program